

Alexander Kopp

List of Publications by Year in Descending Order

Source: <https://exaly.com/author-pdf/1814614/alexander-kopp-publications-by-year.pdf>

Version: 2024-04-28

This document has been generated based on the publications and citations recorded by exaly.com. For the latest version of this publication list, visit the link given above.

The third column is the impact factor (IF) of the journal, and the fourth column is the number of citations of the article.

27
papers

434
citations

11
h-index

20
g-index

30
ext. papers

657
ext. citations

6.2
avg, IF

3.94
L-index

#	Paper	IF	Citations
27	Printability, mechanical and degradation properties of Mg-(x)Zn elemental powder mixes processed by laser powder bed fusion. <i>Additive Manufacturing Letters</i> , 2022 , 2, 100025		0
26	Development and validation of a parametric human mandible model to determine internal stresses for the future design optimization of maxillofacial implants. <i>Journal of the Mechanical Behavior of Biomedical Materials</i> , 2022 , 125, 104893	4.1	0
25	Automated ex-situ detection of pitting corrosion and its effect on the mechanical integrity of rare earth magnesium alloy - WE43. <i>Bioactive Materials</i> , 2022 , 8, 545-558	16.7	4
24	Silk Fibroin as Adjuvant in the Fabrication of Mechanically Stable Fibrin Biocomposites. <i>Polymers</i> , 2022 , 14, 2251	4.5	
23	Influence of surface condition on the degradation behaviour and biocompatibility of additively manufactured WE43. <i>Materials Science and Engineering C</i> , 2021 , 124, 112016	8.3	5
22	Towards a flexible electrochemical biosensor fabricated from biocompatible Bombyx mori silk. <i>Biosensors and Bioelectronics</i> , 2021 , 183, 113204	11.8	5
21	Microstructure, mechanical properties, corrosion resistance and cytocompatibility of WE43 Mg alloy scaffolds fabricated by laser powder bed fusion for biomedical applications. <i>Materials Science and Engineering C</i> , 2021 , 119, 111623	8.3	20
20	Correlates of non-adherence to breast, cervical, and colorectal cancer screening among screen-eligible women: a population-based cohort study in Ontario, Canada. <i>Cancer Causes and Control</i> , 2021 , 32, 147-155	2.8	2
19	Improved in vivo osseointegration and degradation behavior of PEO surface-modified WE43 magnesium plates and screws after 6 and 12 months. <i>Materials Science and Engineering C</i> , 2021 , 129, 112380	8.3	5
18	Strength, corrosion resistance and cellular response of interfaces in bioresorbable poly-lactic acid/Mg fiber composites for orthopedic applications. <i>Journal of the Mechanical Behavior of Biomedical Materials</i> , 2021 , 123, 104781	4.1	3
17	Effect of process parameters on additive-free electrospinning of regenerated silk fibroin nonwovens. <i>Bioactive Materials</i> , 2020 , 5, 241-252	16.7	21
16	A detailed microstructural and corrosion analysis of magnesium alloy WE43 manufactured by selective laser melting. <i>Additive Manufacturing</i> , 2020 , 35, 101321	6.1	14
15	In vitro study on the hemocompatibility of plasma electrolytic oxidation coatings on titanium substrates. <i>Artificial Organs</i> , 2020 , 44, 419-427	2.6	3
14	Introduction to Plasma Electrolytic Oxidation—An Overview of the Process and Applications. <i>Coatings</i> , 2020 , 10, 628	2.9	67
13	In Vivo Simulation of Magnesium Degradability Using a New Fluid Dynamic Bench Testing Approach. <i>International Journal of Molecular Sciences</i> , 2019 , 20,	6.3	9
12	Improved In Vitro Test Procedure for Full Assessment of the Cytocompatibility of Degradable Magnesium Based on ISO 10993-5/-12. <i>International Journal of Molecular Sciences</i> , 2019 , 20,	6.3	32
11	Surface Characteristics of Esthetic Nickel-Titanium and Beta-Titanium Orthodontic Archwires Produced by Plasma Electrolytic Oxidation (PEO)—Primary Results. <i>Materials</i> , 2019 , 12,	3.5	3

10	Defined surface adjustment for medical magnesium implants by electrical discharge machining (EDM) and plasma electrolytic oxidation (PEO). <i>CIRP Annals - Manufacturing Technology</i> , 2019 , 68, 583-586	4.9	4
9	Production and Characterization of Porous Fibroin Scaffolds for Regenerative Medical Application. <i>In Vivo</i> , 2019 , 33, 757-762	2.3	3
8	Influence of design and postprocessing parameters on the degradation behavior and mechanical properties of additively manufactured magnesium scaffolds. <i>Acta Biomaterialia</i> , 2019 , 98, 23-35	10.8	54
7	Hemocompatibility of plasma electrolytic oxidation (PEO) coated Mg-RE and Mg-Zn-Ca alloys for vascular scaffold applications. <i>Materials Science and Engineering C</i> , 2018 , 92, 819-826	8.3	12
6	Plasma Electrolytic Oxidation of Titanium Implant Surfaces: Microgroove-Structures Improve Cellular Adhesion and Viability. <i>In Vivo</i> , 2018 , 32, 241-247	2.3	8
5	PEO-generated Surfaces Support Attachment and Growth of Cells In Vitro with No Additional Benefit for Micro-roughness in Sa (0.2-4 h). <i>In Vivo</i> , 2016 , 30, 27-33	2.3	14
4	Optimized in vitro procedure for assessing the cytocompatibility of magnesium-based biomaterials. <i>Acta Biomaterialia</i> , 2015 , 23, 354-363	10.8	56
3	Influence of Electro Discharge Machining of Biodegradable Magnesium on the Biocompatibility. <i>Procedia CIRP</i> , 2013 , 5, 88-93	1.8	31
2	Multi-scale directed surface topography machined by electro discharge machining in combination with plasma electrolytic conversion for improved osseointegration. <i>International Journal of Mechatronics and Manufacturing Systems</i> , 2013 , 6, 254	0.8	4
1	EDM Machining Capabilities of Magnesium (Mg) Alloy WE43 for Medical Applications. <i>Procedia Engineering</i> , 2011 , 19, 190-195		50