

Marçal Salvadó³ Artells

List of Publications by Year in descending order

Source: <https://exaly.com/author-pdf/6892956/publications.pdf>

Version: 2024-02-01

22
papers

603
citations

759233

12
h-index

713466

21
g-index

23
all docs

23
docs citations

23
times ranked

547
citing authors

#	ARTICLE	IF	CITATIONS
1	Set up of a gamma spectrometry mobile unit equipped with LaBr ₃ (Ce) detectors for radioactivity monitoring. Radiation Physics and Chemistry, 2020, 168, 108600.	2.8	4
2	Calibration and performance of a real-time gamma-ray spectrometry water monitor using a LaBr ₃ (Ce) detector. Radiation Physics and Chemistry, 2018, 144, 444-450.	2.8	4
3	Spectral windows analysis method for monitoring anthropogenic radionuclides in real-time environmental gamma-ray scintillation spectrometry. Journal of Radiological Protection, 2018, 38, 229-246.	1.1	3
4	SimDoseCT: dose reporting software based on Monte Carlo simulation for a 320 detector-row cone-beam CT scanner and ICRP computational adult phantoms. Physics in Medicine and Biology, 2017, 62, 6304-6321.	3.0	11
5	Calculation of the ambient dose equivalent H*(10) from gamma-ray spectra obtained with scintillation detectors. Applied Radiation and Isotopes, 2016, 118, 154-159.	1.5	31
6	Perfusion CT of the Brain and Liver and of Lung Tumors: Use of Monte Carlo Simulation for Patient Dose Estimation for Examinations With a Cone-Beam 320-MDCT Scanner. American Journal of Roentgenology, 2016, 206, 129-135.	2.2	10
7	Monte Carlo simulation of the dose distribution of ICRP adult reference computational phantoms for acquisitions with a 320 detector-row cone-beam CT scanner. Physica Medica, 2015, 31, 452-462.	0.7	7
8	A Monte Carlo simulation for the estimation of patient dose in rest and stress cardiac computed tomography with a 320-detector row CT scanner. Physica Medica, 2015, 31, 1029-1034.	0.7	5
9	Development and calibration of a real-time airborne radioactivity monitor using direct gamma-ray spectrometry with two scintillation detectors. Applied Radiation and Isotopes, 2014, 89, 102-108.	1.5	23
10	Development and Calibration of a Real-Time Airborne Radioactivity Monitor Using Gamma-Ray Spectrometry on a Particulate Filter. IEEE Transactions on Nuclear Science, 2014, 61, 727-731.	2.0	10
11	Implementation of gamma-ray spectrometry in two real-time water monitors using NaI(Tl) scintillation detectors. Applied Radiation and Isotopes, 2013, 80, 49-55.	1.5	34
12	Dosimetry of a cone beam CT device for oral and maxillofacial radiology using Monte Carlo techniques and ICRP adult reference computational phantoms. Dentomaxillofacial Radiology, 2013, 42, 925-933.	2.7	56
13	Temperature peak-shift correction methods for NaI(Tl) and LaBr ₃ (Ce) gamma-ray spectrum stabilisation. Radiation Measurements, 2012, 47, 588-595.	1.4	52
14	Development and calibration of a real-time airborne radioactivity monitor using gamma-ray spectrometry on a particulate filter. , 2012, , .		1
15	Validation of a Monte Carlo simulation for dose assessment in dental cone beam CT examinations. Physica Medica, 2012, 28, 200-209.	0.7	22
16	Energy and resolution calibration of NaI(Tl) and LaBr ₃ (Ce) scintillators and validation of an EGS5 Monte Carlo user code for efficiency calculations. Nuclear Instruments and Methods in Physics Research, Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2012, 675, 78-83.	1.6	61
17	Performance of data acceptance criteria over 50 months from an automatic real-time environmental radiation surveillance network. Journal of Environmental Radioactivity, 2011, 102, 742-748.	1.7	15
18	Radiation Exposure to Patients in a Multicenter Coronary Angiography Trial (CORE 64). American Journal of Roentgenology, 2011, 196, 1126-1132.	2.2	30

#	ARTICLE	IF	CITATIONS
19	Objective assessment of low contrast detectability for real CT phantom and in simulated images using a model observer. , 2011, , .		2
20	Computed tomography dose assessment for a 160 mm wide, 320 detector row, cone beam CT scanner. Physics in Medicine and Biology, 2009, 54, 3141-3159.	3.0	83
21	Quantitative assessment of selective in-plane shielding of tissues in computed tomography through evaluation of absorbed dose and image quality. European Radiology, 2006, 16, 2334-2340.	4.5	111
22	Monte Carlo calculation of radiation dose in CT examinations using phantom and patient tomographic models. Radiation Protection Dosimetry, 2005, 114, 364-368.	0.8	25