

CITATION REPORT

List of articles citing

Fifteen-Year Results of Total Hip Arthroplasty With Cobalt-Chromium Femoral Heads on Highly Cross-Linked Polyethylene in Patients 50 Years and Less

DOI: 10.1016/j.arth.2019.01.071

Journal of Arthroplasty, 2019, 34, 1143-1149.

Source: <https://exaly.com/paper-pdf/73679935/citation-report.pdf>

Version: 2024-04-17

This report has been generated based on the citations recorded by exaly.com for the above article. For the latest version of this publication list, visit the link given above.

The third column is the impact factor (IF) of the journal, and the fourth column is the number of citations of the article.

| # | Paper | IF | Citations |
|----|--|-----|-----------|
| 31 | What's New in Hip Replacement. <i>Journal of Bone and Joint Surgery - Series A</i> , 2019 , 101, 1619-1627 | 5.6 | 9 |
| 30 | The effect of femoral head size on edge loading in metal-on-metal hip joint replacement under dynamic separation conditions. <i>Journal of Biomedical Materials Research - Part B Applied Biomaterials</i> , 2020 , 108, 1897-1906 | 3.5 | 0 |
| 29 | Can the Choice of Cementless Implants and Bearings during Total Hip Arthroplasty Have an Impact on the Overall Costs within a Bundled Payment Model?. <i>The Journal of Hip Surgery</i> , 2020 , 04, 066-076 | 0.2 | |
| 28 | Hip resurfacing arthroplasty in young patients: international high-volume centres' report on the outcome of 11,382 metal-on-metal hip resurfacing arthroplasties in patients >50 years at surgery. <i>HIP International</i> , 2020 , 1120700020957354 | 1.7 | 9 |
| 27 | Recent update on crosslinked polyethylene in total hip arthroplasty. <i>Sicot-j</i> , 2020 , 6, 13 | 2.2 | 9 |
| 26 | In Vitro Wear Performance of X-ray Cross-Linked Vitamin E Blended Polyethylene. <i>Biotribology</i> , 2020 , 21, 100115 | 2.3 | 2 |
| 25 | Incidence and Characteristics of Osteolysis in HXLPE THA at 16-Year Follow up in Patients 50 Years and Less. <i>Journal of Arthroplasty</i> , 2021 , 36, 641-646 | 4.4 | 5 |
| 24 | Polyethylene in Orthopedic Implants: Recent Trends and Limitations. 2021 , | | |
| 23 | A predictive framework of the tribological impact of physical activities on metal-on-plastic hip implants. <i>Biotribology</i> , 2021 , 25, 100156 | 2.3 | 3 |
| 22 | Second-Generation Highly Cross-Linked Polyethylene in Total Hip Arthroplasty. <i>JBJS Reviews</i> , 2021 , 9, e20.00065 | 2.6 | |
| 21 | Highly cross-linked polyethylene is the new gold standard bearing material for total hip arthroplasty. <i>Biosurface and Biotribology</i> , 2021 , 7, 53-59 | 1 | 1 |
| 20 | Treatment of aseptic osteonecrosis of the femoral head: Historical aspects. <i>Morphologie</i> , 2021 , 105, 102-119 | 1.9 | 2 |
| 19 | Total hip arthroplasty using highly cross-linked polyethylene in patients aged 50 years and younger : minimum 15-year follow-up. <i>Bone and Joint Journal</i> , 2021 , 103-B, 78-83 | 5.6 | 3 |
| 18 | Highly Crosslinked Polyethylene Can Reduce Wear Rate in THA for High-Demand Patients: A Matched-Paired Controlled Study. <i>Journal of Arthroplasty</i> , 2021 , 36, 3226-3232 | 4.4 | 0 |
| 17 | Sequentially Irradiated and Annealed Highly Cross-Linked Polyethylene: Linear Vector and Volumetric Wear in Total Hip Arthroplasty at 10 Years. <i>Arthroplasty Today</i> , 2021 , 11, 140-145 | 2 | 1 |
| 16 | Highly Crosslinked Polyethylene Liners Have Negligible Wear at 10 Years: A Radiostereometric Analysis Study. <i>Clinical Orthopaedics and Related Research</i> , 2021 , | 2.2 | 1 |
| 15 | Enhanced computational modelling of UHMWPE wear in total hip joint replacements: The role of frictional work and contact pressure. <i>Wear</i> , 2021 , 482-483, 203985 | 3.5 | 1 |

| | | | |
|----|--|-----|---|
| 14 | CORR Insights : Highly Crosslinked Polyethylene Liners Have Negligible Wear at 10 Years: A Radiostereometric Analysis Study. <i>Clinical Orthopaedics and Related Research</i> , 2021 , | 2.2 | |
| 13 | Wear analysis of the first-generation cross-linked polyethylene at minimum 10 years follow-up after THA: no significant effect of sports participation. <i>Journal of Artificial Organs</i> , 2021 , 1 | 1.8 | 0 |
| 12 | Wear resistance of first-generation highly cross-linked annealed polyethylene in cementless total hip arthroplasty is maintained 20 years after surgery.. <i>Bone and Joint Journal</i> , 2022 , 104-B, 200-205 | 5.6 | 2 |
| 11 | Low Wear at 10-Year Follow-Up of a Second-Generation Highly Cross-Linked Polyethylene in Total Hip Arthroplasty.. <i>Journal of Arthroplasty</i> , 2022 , | 4.4 | 1 |
| 10 | Fixation vs Arthroplasty for Femoral Neck Fracture in Patients Aged 40-59 Years: A Propensity-Score-Matched Analysis.. <i>Arthroplasty Today</i> , 2022 , 14, 175-182 | 2 | |
| 9 | Long-term Follow-up of Conventional Polyethylene in Total Hip Arthroplasty in Young Patients: Heightened Wear-related Complications are Observed at the Beginning of the Third Decade.. <i>Journal of Arthroplasty</i> , 2022 , | 4.4 | |
| 8 | Comparison of modern periacetabular osteotomy for hip dysplasia with total hip arthroplasty for hip osteoarthritis: 10-year outcomes are comparable in young adult patients. <i>Journal of Hip Preservation Surgery</i> , | 2 | 0 |
| 7 | Le PE hautement réticulé dans les prothèses de hanche : 20 ans après. 2022 , | | |
| 6 | Ceramic Femoral Heads Exhibit Lower Wear Rates Compared to Cobalt Chrome: A Meta-Analysis. 2022 , | | 1 |
| 5 | What have we learned from 20 years of using highly crosslinked PE in total hip arthroplasty?. 2022 , 103457 | | 1 |
| 4 | Five-year polyethylene cup migration and PE wear of the Anatomic Dual Mobility acetabular construct. | | 0 |
| 3 | Highly Crosslinked Polyethylene Liner Thickness Does Not Influence Survival in Primary Total Hip Arthroplasty After Mean Follow-Up of 13 Years: A Study of 2,565 Hips With a Single Design Liner. 2023 , | | 0 |
| 2 | Comparative Survival of Contemporary Cementless Acetabular Components Following Revision Total Hip Arthroplasty. 2023 , | | 0 |
| 1 | Mean 16-Year Results of Total Hip Arthroplasty With Alumina Ceramic Femoral Heads on Highly Cross-Linked Polyethylene in Patients 50 Years or Less. 2023 , | | 0 |